

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040119\
 Data File : VN054866.D
 Acq On : 1 Apr 2019 11:33
 Operator : JC/SP
 Sample : VN0401WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampled :
 VN0401WBL01

Quant Time: Apr 02 04:37:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	490502	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	798809	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	712192	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	275224	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	327361	49.51	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.02%	
35) Dibromofluoromethane	7.59	113	254501	47.79	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.58%	
50) Toluene-d8	10.09	98	987848	49.89	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.78%	
62) 4-Bromofluorobenzene	12.40	95	318448	48.36	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.72%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040119\
 Data File : VN054866.D
 Acq On : 1 Apr 2019 11:33
 Operator : JC/SP
 Sample : VN0401WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0401WBL01

Quant Time: Apr 02 04:37:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

