

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040119\
 Data File : VN054873.D
 Acq On : 1 Apr 2019 14:37
 Operator : JC/SP
 Sample : K2178-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-17

Manual Integrations
 APPROVED

MMDadoda
 4/2/2019 10:51:33 AM

Quant Time: Apr 02 05:23:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	531538	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	876683	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	749880	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	263594	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	356541	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
35) Dibromofluoromethane	7.59	113	264316	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	
50) Toluene-d8	10.09	98	1058575	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
62) 4-Bromofluorobenzene	12.40	95	325333	45.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.04%	
Target Compounds						
16) Acetone	3.82	43	78457	45.917	ug/l #	83
20) Methylene Chloride	4.53	84	31274304m	5188.898	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040119\
Data File : VN054873.D
Acq On : 1 Apr 2019 14:37
Operator : JC/SP
Sample : K2178-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-17

Manual Integrations
APPROVED

MMDadoda
4/2/2019 10:51:33 AM

Quant Time: Apr 02 05:23:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

