

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040119\
 Data File : VN054875.D
 Acq On : 1 Apr 2019 15:30
 Operator : JC/SP
 Sample : K2182-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S8A(10)

Quant Time: Apr 02 05:27:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	536523	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	862902	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	751400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	249102	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	349536	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
35) Dibromofluoromethane	7.59	113	280071	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
50) Toluene-d8	10.09	98	1061207	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
62) 4-Bromofluorobenzene	12.40	95	313500	44.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.14%	
Target Compounds						
20) Methylene Chloride	4.55	84	26122	4.294	ug/l #	80
43) Isopropyl Acetate	8.17	43	13989	1.460	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040119\
Data File : VN054875.D
Acq On : 1 Apr 2019 15:30
Operator : JC/SP
Sample : K2182-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S8A(10)

Quant Time: Apr 02 05:27:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

