

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040119\
 Data File : VN054876.D
 Acq On : 1 Apr 2019 15:56
 Operator : JC/SP
 Sample : K2182-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S8A(12)

Quant Time: Apr 02 05:29:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	518422	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	844326	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	735371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	251177	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	342796	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
35) Dibromofluoromethane	7.59	113	271111	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
50) Toluene-d8	10.09	98	1034584	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	12.40	95	314143	45.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.28%	
Target Compounds						
20) Methylene Chloride	4.55	84	21817	3.711	ug/l #	73
43) Isopropyl Acetate	8.17	43	14818	1.581	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040119\
Data File : VN054876.D
Acq On : 1 Apr 2019 15:56
Operator : JC/SP
Sample : K2182-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S8A(12)

Quant Time: Apr 02 05:29:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

