

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040119\
 Data File : VN054883.D
 Acq On : 1 Apr 2019 19:01
 Operator : JC/SP
 Sample : K2186-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BELT-PARKWAY-STAGING-YARD-1

Quant Time: Apr 02 06:54:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	532219	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	865272	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	738777	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	245460	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	344126	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
35) Dibromofluoromethane	7.59	113	272225	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
50) Toluene-d8	10.09	98	1041636	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
62) 4-Bromofluorobenzene	12.40	95	312423	43.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.60%	
Target Compounds						
20) Methylene Chloride	4.54	84	18449	3.057	ug/l	95
43) Isopropyl Acetate	8.17	43	14279	1.486	ug/l #	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040119\
Data File : VN054883.D
Acq On : 1 Apr 2019 19:01
Operator : JC/SP
Sample : K2186-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BELT-PARKWAY-STAGING-YARD-1

Quant Time: Apr 02 06:54:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN054883.D

Acq: 1 Apr 2019 19:01

Instrument :

MSVOA_N

ClientSampleId :


BELT-PARKWAY-STAGING-YARD-1

Tot Ion:168 Resp: 532219

Ion Ratio Lower Upper

168 100

99 54.9 37.1 55.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

250000

200000



150000

100000

50000

0

Time-->



#20

Methylene Chloride

Concen: 3.057 ug/l

RT: 4.54 min Scan# 862

Delta R.T. -0.00 min

Lab File: VN054883.D

Acq: 1 Apr 2019 19:01

Tgt Ion: 84 Resp: 18449


Ion Ratio Lower Upper

84 100

49 115.3 89.7 134.5

51 36.7 28.2 42.4

86 55.1 50.8 76.2



Abundance Ion 83.90 (83.60 to 84.60): VNO

Ion 48.90 (48.60 to 49.60): VNO

Ion 50.90 (50.60 to 51.60): VNO

Ion 85.90 (85.60 to 86.60): VNO

4.54

12000

10000

8000


6000


4000

2000

0

Time-->









#33
 1,2-Dichloroethane-d4
 Concen: 47.966 ug/l
 RT: 8.03 min Scan# 1945
 Delta R.T. -0.00 min
 Lab File: VN054883.D
 Acq: 1 Apr 2019 19:01




Instrument :
 MSVOA_N
 ClientSampleId :
 BELT-PARKWAY-STAGING-YARD-1


Tot Ion: 65 Resp: 344126
 Ion Ratio Lower Upper
 65 100
 67 52.7 0.0 104.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. -0.00 min
 Lab File: VN054883.D
 Acq: 1 Apr 2019 19:01

Tgt Ion: 114 Resp: 865272
 Ion Ratio Lower Upper
 114 100
 63 22.2 0.0 35.2
 88 15.9 0.0 30.2





#35

Dibromofluoromethane

Concen: 47.190 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VN054883.D

Acq: 1 Apr 2019 19:01

Instrument :

MSVOA_N

ClientSampleId :

BELT-PARKWAY-STAGING-YARD-1


Tot Ion:113 Resp: 272225

Ion Ratio Lower Upper

113 100

111 103.2 82.2 123.4




192 20.6 19.2 28.8



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#43

Isopropyl Acetate

Concen: 1.486 ug/l

RT: 8.17 min Scan# 1988

Delta R.T. -0.00 min

Lab File: VN054883.D

Acq: 1 Apr 2019 19:01


Tgt Ion: 43 Resp: 14279

Ion Ratio Lower Upper

43 100

61 17.3 15.8 23.8



87 7.3 12.0 18.0#




Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 61.00 (60.70 to 61.70): VNO

Ion 87.00 (86.70 to 87.70): VNO





#50

Toluene-d8

Concen: 48.568 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN054883.D

Acq: 1 Apr 2019 19:01

Instrument :

MSVOA_N

ClientSampleId :


BELT-PARKWAY-STAGING-YARD-1

Tot Ion: 98 Resp: 1041636

Ion Ratio Lower Upper

98 100

100 64.1 51.2 76.8



Abundance Ion 98.00 (97.70 to 98.70): VN054762.D (-2569) (-)

Ion 100.00 (99.70 to 100.70): VN054762.D (-2569) (-)

10.09

600000

500000

400000



300000

200000

100000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 43.802 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN054883.D

Acq: 1 Apr 2019 19:01


Tgt Ion: 95 Resp: 312423

Ion Ratio Lower Upper

95 100

174 84.0 0.0 191.0

176 81.7 0.0 187.4



Abundance Ion 94.90 (94.60 to 95.60): VN054762.D (-3291) (-)

Ion 173.80 (173.50 to 174.50): VN054762.D (-3291) (-)

Ion 175.80 (175.50 to 176.50): VN054762.D (-3291) (-)

12.40

200000


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN054883.D
Acq: 1 Apr 2019 19:01

Instrument :
MSVOA_N
ClientSampleId :
BELT-PARKWAY-STAGING-YARD-1

Tot Ion:117 Resp: 738777
Ion Ratio Lower Upper
117 100
82 55.5 40.0 60.0
119 31.5 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. -0.00 min
Lab File: VN054883.D
Acq: 1 Apr 2019 19:01

Tgt Ion:152 Resp: 245460
Ion Ratio Lower Upper
152 100
115 57.7 27.2 81.6
150 156.6 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

