


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040119\
Data File : VN054864.D
Acq On : 1 Apr 2019 9:29
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Title : SW846 8260
Last Update : Sat Mar 30 00:48:18 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	11775	PASS
75	95	30	60	52.9	28986	PASS
95	95	100	100	100.0	54765	PASS
96	95	5	9	6.7	3665	PASS
173	174	0.00	2	2.0	922	PASS
174	95	50	100	85.7	46954	PASS
175	174	5	9	8.3	3888	PASS
176	174	95	101	95.9	45037	PASS
177	176	5	9	5.9	2635	PASS