

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040120\
 Data File : VN060810.D
 Acq On : 1 Apr 2020 15:03
 Operator : JC/MD
 Sample : L2106-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Z4-EB-200331

Quant Time: Apr 02 09:24:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	191456	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	318495	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	292498	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	128635	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	123860	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
35) Dibromofluoromethane	7.56	113	94348	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
50) Toluene-d8	10.08	98	387619	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
62) 4-Bromofluorobenzene	12.40	95	142489	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
Target Compounds						
44) Trichloroethene	8.82	130	10038	4.277	ug/l	97
64) Tetrachloroethene	10.62	164	4919	2.137	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040120\
Data File : VN060810.D
Acq On : 1 Apr 2020 15:03
Operator : JC/MD
Sample : L2106-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Z4-EB-200331

Quant Time: Apr 02 09:24:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

