

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040121\
 Data File : VN066456.D
 Acq On : 1 Apr 2021 15:24
 Operator : JC/MD
 Sample : M1836-06
 Misc : 10.09g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B3-10-12

Quant Time: Apr 02 04:51:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 01 14:08:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.579	168	459635	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	672273	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	533310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	209706	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.944	65	276693	45.28	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.56%
35) Dibromofluoromethane	7.504	113	226446	50.81	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.62%
50) Toluene-d8	10.025	98	826440	48.89	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.78%
62) 4-Bromofluorobenzene	12.345	95	235484	40.11	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.22%#
Target Compounds						
						Qvalue
16) Acetone	3.738	43	8257	2.50	ug/l	91
18) Methyl Acetate	4.235	43	52707	6.11	ug/l	96
95) Naphthalene	15.268	128	43448	4.66	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040121\
Data File : VN066456.D
Acq On : 1 Apr 2021 15:24
Operator : JC/MD
Sample : M1836-06
Misc : 10.09g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B3-10-12

Quant Time: Apr 02 04:51:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 01 14:08:22 2021
Response via : Initial Calibration

