

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040121\
 Data File : VN066448.D
 Acq On : 1 Apr 2021 11:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 04:47:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 01 14:08:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.630	0.648	-2.9	82	0.00
3 P	Chloromethane	0.739	0.702	5.0	85	0.00
4 C	Vinyl Chloride	0.721	0.767	-6.4#	87	0.00
5 T	Bromomethane	0.483	0.523	-8.3	86	0.00
6 T	Chloroethane	0.449	0.480	-6.9	87	0.00
7 T	Trichlorofluoromethane	0.991	1.025	-3.4	83	0.00
8 T	Diethyl Ether	0.404	0.422	-4.5	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.560	0.588	-5.0	85	0.00
10 T	Methyl Iodide	0.794	0.688	13.4	67	0.00
11 T	Tert butyl alcohol	0.101	0.108	-6.9	80	0.00
12 CM	1,1-Dichloroethene	0.573	0.570	0.5#	81	0.00
13 T	Acrolein	0.066	0.041	37.9#	61	0.00
14 T	Allyl chloride	1.133	1.257	-10.9	87	0.00
15 T	Acrylonitrile	0.283	0.303	-7.1	84	0.00
16 T	Acetone	0.359	0.403	-12.3	92	0.00
17 T	Carbon Disulfide	1.692	1.689	0.2	82	0.00
18 T	Methyl Acetate	0.938	1.165	-24.2	98	0.00
19 T	Methyl tert-butyl Ether	1.699	1.637	3.6	79	0.00
20 T	Methylene Chloride	0.692	0.696	-0.6	86	0.00
21 T	trans-1,2-Dichloroethene	0.520	0.498	4.2	79	0.00
22 T	Diisopropyl ether	1.768	1.826	-3.3	84	0.00
23 T	Vinyl Acetate	1.505	1.567	-4.1	82	0.00
24 P	1,1-Dichloroethane	0.976	0.987	-1.1	82	0.00
25 T	2-Butanone	0.393	0.432	-9.9	86	0.00
26 T	2,2-Dichloropropane	0.870	0.837	3.8	78	0.00
27 T	cis-1,2-Dichloroethene	0.606	0.588	3.0	80	0.00
28 T	Bromochloromethane	0.492	0.486	1.2	80	0.00
29 T	Tetrahydrofuran	0.256	0.286	-11.7	86	0.00
30 C	Chloroform	0.976	0.991	-1.5#	83	0.00
31 T	Cyclohexane	0.920	0.852	7.4	78	0.00
32 T	1,1,1-Trichloroethane	0.857	0.834	2.7	78	0.00
33 S	1,2-Dichloroethane-d4	0.665	0.626	5.9	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.331	0.319	3.6	77	0.00
36 T	1,1-Dichloropropene	0.470	0.465	1.1	80	0.00
37 T	Ethyl Acetate	0.517	0.575	-11.2	86	0.00
38 T	Carbon Tetrachloride	0.472	0.460	2.5	78	0.00
39 T	Methylcyclohexane	0.573	0.554	3.3	77	0.00
40 TM	Benzene	1.419	1.448	-2.0	81	0.00
41 T	Methacrylonitrile	0.257	0.268	-4.3	88	0.00
42 TM	1,2-Dichloroethane	0.508	0.519	-2.2	82	0.00
43 T	Isopropyl Acetate	0.884	0.938	-6.1	83	0.00
44 TM	Trichloroethene	0.373	0.362	2.9	78	0.00
45 C	1,2-Dichloropropane	0.378	0.398	-5.3#	84	0.00
46 T	Dibromomethane	0.248	0.256	-3.2	81	0.00
47 T	Bromodichloromethane	0.512	0.526	-2.7	82	0.00
48 T	Methyl methacrylate	0.420	0.434	-3.3	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040121\
 Data File : VN066448.D
 Acq On : 1 Apr 2021 11:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 04:47:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 01 14:08:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.006	-20.0	83	0.00
50 S	Toluene-d8	1.257	1.188	5.5	75	0.00
51 T	4-Methyl-2-Pentanone	0.516	0.584	-13.2	86	0.00
52 CM	Toluene	0.871	0.900	-3.3#	81	0.00
53 T	t-1,3-Dichloropropene	0.556	0.578	-4.0	82	0.00
54 T	cis-1,3-Dichloropropene	0.597	0.616	-3.2	81	0.00
55 T	1,1,2-Trichloroethane	0.350	0.371	-6.0	83	0.00
56 T	Ethyl methacrylate	0.537	0.558	-3.9	80	0.00
57 T	1,3-Dichloropropane	0.581	0.613	-5.5	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.048	0.045	6.3	71	0.00
59 T	2-Hexanone	0.362	0.422	-16.6	85	0.00
60 T	Dibromochloromethane	0.389	0.408	-4.9	81	0.00
61 T	1,2-Dibromoethane	0.365	0.379	-3.8	81	0.00
62 S	4-Bromofluorobenzene	0.437	0.422	3.4	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.426	0.382	10.3	74	0.00
65 PM	Chlorobenzene	0.997	1.007	-1.0	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.384	0.384	0.0	80	0.00
67 C	Ethyl Benzene	1.809	1.819	-0.6#	80	0.00
68 T	m/p-Xylenes	0.688	0.684	0.6	79	0.00
69 T	o-Xylene	0.663	0.651	1.8	78	0.00
70 T	Styrene	1.076	1.117	-3.8	80	0.00
71 P	Bromoform	0.319	0.326	-2.2	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.719	3.364	9.5	79	0.00
74 T	N-amyl acetate	1.307	1.317	-0.8	77	0.00
75 P	1,1,2,2-Tetrachloroethane	1.184	1.160	2.0	87	0.00
76 T	1,2,3-Trichloropropane	1.097	1.129	-2.9	90	0.00
77 T	Bromobenzene	0.939	0.867	7.7	81	0.00
78 T	n-propylbenzene	4.044	3.929	2.8	82	0.00
79 T	2-Chlorotoluene	2.487	2.267	8.8	80	0.00
80 T	1,3,5-Trimethylbenzene	3.042	2.769	9.0	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.365	0.367	-0.5	81	0.00
82 T	4-Chlorotoluene	2.475	2.356	4.8	80	0.00
83 T	tert-Butylbenzene	2.619	2.303	12.1	77	0.00
84 T	1,2,4-Trimethylbenzene	2.956	2.722	7.9	78	0.00
85 T	sec-Butylbenzene	3.393	3.223	5.0	79	0.00
86 T	p-Isopropyltoluene	3.039	2.869	5.6	78	0.00
87 T	1,3-Dichlorobenzene	1.576	1.521	3.5	82	0.00
88 T	1,4-Dichlorobenzene	1.636	1.545	5.6	81	0.00
89 T	n-Butylbenzene	2.510	2.509	0.0	82	0.00
90 T	Hexachloroethane	0.548	0.515	6.0	81	0.00
91 T	1,2-Dichlorobenzene	1.549	1.497	3.4	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.227	0.227	0.0	82	0.00
93 T	1,2,4-Trichlorobenzene	0.961	0.945	1.7	83	0.00
94 T	Hexachlorobutadiene	0.524	0.498	5.0	82	0.00
95 T	Naphthalene	2.899	2.636	9.1	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040121\
Data File : VN066448.D
Acq On : 1 Apr 2021 11:41
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 02 04:47:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 01 14:08:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.966	0.912	5.6	83	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6