

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071748.D
 Acq On : 01 Apr 2022 13:18
 Operator : JC\MD
 Sample : N2075-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT161S1-20220321

Quant Time: Apr 02 00:52:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/02/2022
 Supervised By : Mahesh Dadoda 04/02/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	539568	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	922330	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	897411	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	295816	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	308189	50.032	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.060%			
35) Dibromofluoromethane	8.016	113	263112	48.479	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.960%			
50) Toluene-d8	10.439	98	1149293	51.765	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.520%			
62) 4-Bromofluorobenzene	12.727	95	346014	48.188	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 96.380%			
Target Compounds					Qvalue	
16) Acetone	4.287	43	8873m	3.371	ug/l	
24) 1,1-Dichloroethane	6.387	63	27536	2.981	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071748.D
Acq On : 01 Apr 2022 13:18
Operator : JC\MD
Sample : N2075-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT161S1-20220321

Manual Integrations
APPROVED

Reviewed By : John Carlone 04/02/2022
Supervised By : Mahesh Dadoda 04/02/2022

Quant Time: Apr 02 00:52:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

