

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071752.D
 Acq On : 01 Apr 2022 14:55
 Operator : JC\MD
 Sample : N2075-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-178S-20220322

Quant Time: Apr 02 00:54:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	615355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1043439	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1040161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	363196	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	353190	50.276	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.560%	
35) Dibromofluoromethane	8.022	113	300051	48.868	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.740%	
50) Toluene-d8	10.439	98	1301653	51.823	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.640%	
62) 4-Bromofluorobenzene	12.727	95	422530	52.014	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 104.020%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	5569	1.855	ug/l	95
19) Methyl tert-butyl Ether	5.616	73	27113	1.437	ug/l #	91
44) Trichloroethene	9.210	130	2397	0.344	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071752.D
Acq On : 01 Apr 2022 14:55
Operator : JC\MD
Sample : N2075-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-178S-20220322

Quant Time: Apr 02 00:54:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

