

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071753.D
 Acq On : 01 Apr 2022 15:19
 Operator : JC\MD
 Sample : N2090-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-47

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/02/2022
 Supervised By : Mahesh Dadoda 04/02/2022

Quant Time: Apr 02 00:54:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	607191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1043050	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1064153	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	406795	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	356139	51.378	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.760%			
35) Dibromofluoromethane	8.016	113	302622	49.305	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.620%			
50) Toluene-d8	10.439	98	1301029	51.817	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.640%			
62) 4-Bromofluorobenzene	12.727	95	453176	55.807	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 111.620%			
Target Compounds						
11) Tert butyl alcohol	5.363	59	17217	15.305	ug/l	96
16) Acetone	4.293	43	29322	9.898	ug/l	95
19) Methyl tert-butyl Ether	5.610	73	33568	1.803	ug/l	95
25) 2-Butanone	7.351	43	7457m	1.746	ug/l	
40) Benzene	8.457	78	10992	0.406	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071753.D
Acq On : 01 Apr 2022 15:19
Operator : JC\MD
Sample : N2090-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-47

Quant Time: Apr 02 00:54:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/02/2022
Supervised By :Mahesh Dadoda 04/02/2022

