

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071759.D
 Acq On : 01 Apr 2022 17:44
 Operator : JC\MD
 Sample : N2090-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-53

Quant Time: Apr 02 00:56:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	624360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1076827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1095298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	399035	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	358005	50.227	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.460%	
35) Dibromofluoromethane	8.022	113	307085	48.463	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.920%	
50) Toluene-d8	10.439	98	1342180	51.779	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.560%	
62) 4-Bromofluorobenzene	12.727	95	450355	53.720	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.440%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.363	59	532623	460.458	ug/l	99
16) Acetone	4.298	43	11769	3.864	ug/l	96
19) Methyl tert-butyl Ether	5.616	73	472466	24.678	ug/l	96
22) Diisopropyl ether	6.492	45	49519	2.719	ug/l	94
40) Benzene	8.457	78	717962	25.685	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071759.D
Acq On : 01 Apr 2022 17:44
Operator : JC\MD
Sample : N2090-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-53

Quant Time: Apr 02 00:56:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

