

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071765.D
 Acq On : 01 Apr 2022 20:09
 Operator : JC\MD
 Sample : N2090-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-60

Quant Time: Apr 02 00:57:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	649642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1107803	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1125426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	405874	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	375691	50.657	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.320%	
35) Dibromofluoromethane	8.022	113	315078	48.334	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.660%	
50) Toluene-d8	10.439	98	1397409	52.402	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.800%	
62) 4-Bromofluorobenzene	12.727	95	457321	53.026	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 106.060%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.369	59	45176	37.535	ug/l #	84
16) Acetone	4.292	43	15886	5.012	ug/l	92
19) Methyl tert-butyl Ether	5.616	73	203074	10.194	ug/l	99
22) Diisopropyl ether	6.492	45	7133	0.376	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071765.D
Acq On : 01 Apr 2022 20:09
Operator : JC\MD
Sample : N2090-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-60

Quant Time: Apr 02 00:57:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

