

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN040122\  
 Data File : VN071766.D  
 Acq On : 01 Apr 2022 20:33  
 Operator : JC\MD  
 Sample : N2090-19  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-63

Quant Time: Apr 02 00:57:55 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N033022W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Mar 31 01:57:54 2022  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|------------|----------|
| -----                       |           |       |          |          |            |          |
| Internal Standards          |           |       |          |          |            |          |
| 1) Pentafluorobenzene       | 8.086     | 168   | 629082   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.969     | 114   | 1078467  | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 11.745    | 117   | 1113045  | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.674    | 152   | 385946   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |           |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 8.439     | 65    | 364604   | 50.768   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 61 - 141 | Recovery | = 101.540% |          |
| 35) Dibromofluoromethane    | 8.022     | 113   | 313357   | 49.377   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 69 - 133 | Recovery | = 98.760%  |          |
| 50) Toluene-d8              | 10.439    | 98    | 1382321  | 53.247   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 65 - 126 | Recovery | = 106.500% |          |
| 62) 4-Bromofluorobenzene    | 12.727    | 95    | 441342   | 52.565   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 58 - 135 | Recovery | = 105.140% |          |
| Target Compounds            |           |       |          |          |            |          |
|                             |           |       |          |          |            | Qvalue   |
| 16) Acetone                 | 4.298     | 43    | 9501     | 3.096    | ug/l       | 97       |
| 19) Methyl tert-butyl Ether | 5.622     | 73    | 12299    | 0.638    | ug/l       | 99       |
| -----                       |           |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN040122\  
Data File : VN071766.D  
Acq On : 01 Apr 2022 20:33  
Operator : JC\MD  
Sample : N2090-19  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-63

Quant Time: Apr 02 00:57:55 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N033022W.M  
Quant Title : SW846 8260  
QLast Update : Thu Mar 31 01:57:54 2022  
Response via : Initial Calibration

