

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071767.D
 Acq On : 01 Apr 2022 20:57
 Operator : JC\MD
 Sample : N2090-21
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP032322

Quant Time: Apr 02 00:58:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	623206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1058273	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1088844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	406700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	351085	49.347	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.700%
35) Dibromofluoromethane	8.022	113	299500	48.094	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.180%
50) Toluene-d8	10.439	98	1357949	53.306	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.620%
62) 4-Bromofluorobenzene	12.727	95	457665	55.549	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.100%
Target Compounds						
					Qvalue	
25) 2-Butanone	7.334	43	14121	3.222	ug/l #	56
31) Cyclohexane	8.093	56	52340	5.410	ug/l #	83
40) Benzene	8.463	78	16437	0.598	ug/l	96
67) Ethyl Benzene	11.839	91	8912	0.241	ug/l #	77
68) m/p-Xylenes	11.951	106	5100	0.360	ug/l	77
83) tert-Butylbenzene	13.322	119	7162	0.289	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071767.D
Acq On : 01 Apr 2022 20:57
Operator : JC\MD
Sample : N2090-21
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP032322

Quant Time: Apr 02 00:58:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

