

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071771.D
 Acq On : 01 Apr 2022 23:21
 Operator : JC\MD
 Sample : VN0401WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0401WBL02

Quant Time: Apr 02 00:59:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	617943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1053309	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1069259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	368560	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	354189	50.207	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.420%
35) Dibromofluoromethane	8.022	113	298458	48.153	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.300%
50) Toluene-d8	10.439	98	1321023	52.101	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.200%
62) 4-Bromofluorobenzene	12.727	95	428844	52.297	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071771.D
Acq On : 01 Apr 2022 23:21
Operator : JC\MD
Sample : VN0401WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0401WBL02

Quant Time: Apr 02 00:59:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

