

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071783.D
 Acq On : 02 Apr 2022 04:10
 Operator : JC\MD
 Sample : N2222-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-22

Quant Time: Apr 02 05:34:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	629913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1077923	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1067911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	375907	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	358956	49.916	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.840%
35) Dibromofluoromethane	8.016	113	308913	48.702	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.400%
50) Toluene-d8	10.439	98	1333107	51.377	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.760%
62) 4-Bromofluorobenzene	12.727	95	432490	51.537	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.080%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	9151	2.978	ug/l #	79
43) Isopropyl Acetate	8.622	43	3540	0.220	ug/l #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071783.D
Acq On : 02 Apr 2022 04:10
Operator : JC\MD
Sample : N2222-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-22

Quant Time: Apr 02 05:34:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

