

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071784.D
 Acq On : 02 Apr 2022 04:34
 Operator : JC\MD
 Sample : N2222-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Apr 02 05:34:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	614082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1045489	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1044413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	348241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	354968	50.634	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	101.260%	
35) Dibromofluoromethane	8.022	113	300368	48.824	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	97.640%	
50) Toluene-d8	10.439	98	1313717	52.200	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	104.400%	
62) 4-Bromofluorobenzene	12.727	95	408830	50.229	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	100.460%	
Target Compounds						
						Qvalue
16) Acetone	4.304	43	5444	1.817	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071784.D
Acq On : 02 Apr 2022 04:34
Operator : JC\MD
Sample : N2222-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Apr 02 05:34:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

