

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071786.D
 Acq On : 02 Apr 2022 05:23
 Operator : JC\MD
 Sample : N2200-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-26

Quant Time: Apr 02 06:13:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	622995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1062365	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1072722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	370833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	360531	50.692	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.380%
35) Dibromofluoromethane	8.022	113	306737	49.067	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.140%
50) Toluene-d8	10.439	98	1343354	52.530	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.060%
62) 4-Bromofluorobenzene	12.727	95	423798	51.241	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.480%
Target Compounds						
16) Acetone	4.305	43	5476	1.802	ug/l #	81
19) Methyl tert-butyl Ether	5.622	73	377745	19.773	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071786.D
Acq On : 02 Apr 2022 05:23
Operator : JC\MD
Sample : N2200-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-26

Quant Time: Apr 02 06:13:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

