

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071788.D
 Acq On : 02 Apr 2022 06:11
 Operator : JC\MD
 Sample : N2200-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-30D

Quant Time: Apr 02 07:14:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	623701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1074879	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1078090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	383249	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	360802	50.672	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.340%
35) Dibromofluoromethane	8.022	113	306409	48.444	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.880%
50) Toluene-d8	10.439	98	1342517	51.886	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.780%
62) 4-Bromofluorobenzene	12.727	95	432019	51.627	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.260%
Target Compounds						
						Qvalue
16) Acetone	4.299	43	5925	1.947	ug/l	98
19) Methyl tert-butyl Ether	5.610	73	65289	3.414	ug/l	96
40) Benzene	8.463	78	5800	0.208	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071788.D
Acq On : 02 Apr 2022 06:11
Operator : JC\MD
Sample : N2200-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-30D

Quant Time: Apr 02 07:14:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

