

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071789.D
 Acq On : 02 Apr 2022 06:35
 Operator : JC\MD
 Sample : N2200-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-31

Quant Time: Apr 02 07:15:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	613161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1034585	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1054121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	382000	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	355560	50.795	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery =	101.580%		
35) Dibromofluoromethane	8.022	113	301424	49.512	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery =	99.020%		
50) Toluene-d8	10.439	98	1320752	53.033	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery =	106.060%		
62) 4-Bromofluorobenzene	12.727	95	431212	53.537	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery =	107.080%		
Target Compounds						
						Qvalue
16) Acetone	4.287	43	51918	17.355	ug/l	94
17) Carbon Disulfide	4.540	76	82410	7.097	ug/l	100
25) 2-Butanone	7.345	43	23482	5.445	ug/l	90
27) cis-1,2-Dichloroethene	7.328	96	20598	3.023	ug/l	90
44) Trichloroethene	9.210	130	11362	1.642	ug/l	92
51) 4-Methyl-2-Pentanone	10.328	43	61524	6.914	ug/l	100
64) Tetrachloroethene	10.975	164	45711	6.567	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071789.D
 Acq On : 02 Apr 2022 06:35
 Operator : JC\MD
 Sample : N2200-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-31

Quant Time: Apr 02 07:15:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

