

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071790.D
 Acq On : 02 Apr 2022 06:59
 Operator : JC\MD
 Sample : N2200-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-32

Quant Time: Apr 02 07:50:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	593678	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1022102	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1048029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	380109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	347650	51.295	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.580%
35) Dibromofluoromethane	8.022	113	296245	49.255	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.520%
50) Toluene-d8	10.439	98	1287421	52.326	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.660%
62) 4-Bromofluorobenzene	12.727	95	423370	53.205	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.420%
Target Compounds						
						Qvalue
16) Acetone	4.292	43	5200	1.795	ug/l	91
17) Carbon Disulfide	4.545	76	6104	0.543	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071790.D
Acq On : 02 Apr 2022 06:59
Operator : JC\MD
Sample : N2200-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-32

Quant Time: Apr 02 07:50:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

