

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
 Data File : VN071793.D
 Acq On : 02 Apr 2022 08:11
 Operator : JC\MD
 Sample : N2200-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-35

Quant Time: Apr 02 08:52:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	620331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1070389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1091120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	378284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	364534	51.475	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.940%	
35) Dibromofluoromethane	8.022	113	303647	48.208	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.420%	
50) Toluene-d8	10.439	98	1338156	51.935	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.860%	
62) 4-Bromofluorobenzene	12.727	95	447474	53.698	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.400%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.363	59	228236	198.594	ug/l	100
16) Acetone	4.299	43	7303	2.413	ug/l	95
19) Methyl tert-butyl Ether	5.616	73	122170	6.423	ug/l	96
22) Diisopropyl ether	6.498	45	17271	0.955	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040122\
Data File : VN071793.D
Acq On : 02 Apr 2022 08:11
Operator : JC\MD
Sample : N2200-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-35

Quant Time: Apr 02 08:52:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

