

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040124\
 Data File : VN081637.D
 Acq On : 01 Apr 2024 10:13
 Operator : JC\MD
 Sample : VN0401MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0401MBL01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/02/2024
 Supervised By : Mahesh Dadoda 04/02/2024

Quant Time: Apr 02 01:10:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.871	168	368829m	50.000	ug/l	-0.35
34) 1,4-Difluorobenzene	8.859	114	654042	50.000	ug/l	-0.25
63) Chlorobenzene-d5	11.812	117	588436m	50.000	ug/l	-0.05
72) 1,4-Dichlorobenzene-d4	13.782	152	227959	50.000	ug/l	-0.01

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.259	65	247366	52.591	ug/l	-0.32
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.180%	
35) Dibromofluoromethane	7.788	113	215798	52.281	ug/l	-0.38
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.560%	
50) Toluene-d8	10.453	98	783910	52.537	ug/l	-0.12
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.080%	
62) 4-Bromofluorobenzene	12.823	95	236915m	45.061	ug/l	-0.03
Spiked Amount	50.000	Range	64 - 133	Recovery	= 90.120%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040124\
Data File : VN081637.D
Acq On : 01 Apr 2024 10:13
Operator : JC\MD
Sample : VN0401MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0401MBL01

Manual Integrations
APPROVED

Reviewed By : John Carlone 04/02/2024
Supervised By : Mahesh Dadoda 04/02/2024

Quant Time: Apr 02 01:10:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

