

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040124\
 Data File : VN081638.D
 Acq On : 01 Apr 2024 10:36
 Operator : JC\MD
 Sample : VN0401WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0401WBL01

Quant Time: Apr 02 01:10:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	359075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	655996	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	602174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	244192	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	249228	54.427	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.860%
35) Dibromofluoromethane	8.165	113	213782	51.638	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.280%
50) Toluene-d8	10.565	98	785815	52.508	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.020%
62) 4-Bromofluorobenzene	12.847	95	256421	48.626	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	97.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040124\
Data File : VN081638.D
Acq On : 01 Apr 2024 10:36
Operator : JC\MD
Sample : VN0401WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0401WBL01

Quant Time: Apr 02 01:10:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

