

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040125\
 Data File : VN086191.D
 Acq On : 01 Apr 2025 12:30
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Apr 02 01:51:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	131862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	204993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	195897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	111589	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	97838	54.185	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.360%			
35) Dibromofluoromethane	8.165	113	74351	51.964	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.920%			
50) Toluene-d8	10.565	98	283232	54.542	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.080%			
62) 4-Bromofluorobenzene	12.847	95	105921	57.194	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 114.380%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	105974	60.662	ug/l	98
3) Chloromethane	2.359	50	86301	56.350	ug/l	98
4) Vinyl Chloride	2.506	62	91233	58.261	ug/l	99
5) Bromomethane	2.936	94	50856	47.678	ug/l	98
6) Chloroethane	3.106	64	48523	49.119	ug/l	95
7) Trichlorofluoromethane	3.489	101	140135	49.782	ug/l	99
8) Diethyl Ether	3.959	74	36707	42.596	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	70730	47.277	ug/l	99
10) Methyl Iodide	4.583	142	94384	48.417	ug/l	93
11) Tert butyl alcohol	5.524	59	56164	219.694	ug/l	98
12) 1,1-Dichloroethene	4.336	96	63208	46.789	ug/l	96
13) Acrolein	4.177	56	49816	182.699	ug/l	94
14) Allyl chloride	5.018	41	67909	40.294	ug/l	93
15) Acrylonitrile	5.718	53	159244	235.861	ug/l	98
16) Acetone	4.424	43	116296	210.819	ug/l	97
17) Carbon Disulfide	4.712	76	194846	43.857	ug/l	99
18) Methyl Acetate	5.024	43	63507	46.530	ug/l	97
19) Methyl tert-butyl Ether	5.794	73	217945	49.233	ug/l	100
20) Methylene Chloride	5.277	84	77510	48.030	ug/l	96
21) trans-1,2-Dichloroethene	5.788	96	73509	49.791	ug/l	96
22) Diisopropyl ether	6.671	45	172108	47.921	ug/l	99
23) Vinyl Acetate	6.600	43	656603	241.172	ug/l	98
24) 1,1-Dichloroethane	6.565	63	127834	48.404	ug/l	100
25) 2-Butanone	7.482	43	185831	233.461	ug/l	100
26) 2,2-Dichloropropane	7.488	77	120377	47.344	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	85938	51.224	ug/l	98
28) Bromochloromethane	7.812	49	54285	48.515	ug/l	99
29) Tetrahydrofuran	7.835	42	124704	253.767	ug/l	98
30) Chloroform	7.965	83	151272	51.821	ug/l	97
31) Cyclohexane	8.253	56	105197	46.757	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	146295	52.943	ug/l	98
36) 1,1-Dichloropropene	8.371	75	101016	53.617	ug/l	99
37) Ethyl Acetate	7.559	43	77375	48.680	ug/l	100
38) Carbon Tetrachloride	8.359	117	141628	57.173	ug/l	98
39) Methylcyclohexane	9.600	83	96125	51.425	ug/l	90
40) Benzene	8.606	78	317513	53.684	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040125\
 Data File : VN086191.D
 Acq On : 01 Apr 2025 12:30
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Amit Patel 04/02/2025

Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 01:51:52 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M

Quant Title : SW846 8260

QLast Update : Wed Mar 19 03:20:56 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	41220	54.310	ug/l	96
42) 1,2-Dichloroethane	8.665	62	113538	55.093	ug/l	99
43) Isopropyl Acetate	8.688	43	135463	39.330	ug/l	99
44) Trichloroethene	9.353	130	75257	49.095	ug/l	95
45) 1,2-Dichloropropane	9.618	63	70881	52.691	ug/l	92
46) Dibromomethane	9.706	93	58614	55.405	ug/l	99
47) Bromodichloromethane	9.888	83	122368	55.345	ug/l	99
48) Methyl methacrylate	9.676	41	61608	54.205	ug/l	96
49) 1,4-Dioxane	9.694	88	30471	1134.096	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	433441	285.346	ug/l	99
52) Toluene	10.629	92	212069	58.421	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	115814	53.936	ug/l	95
54) cis-1,3-Dichloropropene	10.312	75	121877	53.797	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	77847	56.188	ug/l	96
56) Ethyl methacrylate	10.871	69	111441	51.295	ug/l	98
57) 1,3-Dichloropropane	11.159	76	132157	56.160	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	263133	304.296	ug/l	99
59) 2-Hexanone	11.194	43	321084	290.158	ug/l	99
60) Dibromochloromethane	11.359	129	102372	57.261	ug/l	99
61) 1,2-Dibromoethane	11.470	107	78903	55.473	ug/l	98
64) Tetrachloroethene	11.100	164	80488	51.506	ug/l	97
65) Chlorobenzene	11.888	112	225670	50.343	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	87978	52.796	ug/l	98
67) Ethyl Benzene	11.965	91	386894	53.956	ug/l	99
68) m/p-Xylenes	12.070	106	317096	112.511	ug/l	98
69) o-Xylene	12.394	106	145782	56.358	ug/l	100
70) Styrene	12.412	104	260792	59.027	ug/l	99
71) Bromoform	12.576	173	73884	53.039	ug/l #	99
73) Isopropylbenzene	12.694	105	373293	48.987	ug/l	100
74) N-amyl acetate	12.494	43	119388	46.828	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	114711	44.496	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	113386m	45.386	ug/l	
77) Bromobenzene	12.976	156	100250	47.181	ug/l	97
78) n-propylbenzene	13.035	91	443917	51.446	ug/l	100
79) 2-Chlorotoluene	13.123	91	281383	49.095	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	340956	52.995	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	40484	42.843	ug/l	92
82) 4-Chlorotoluene	13.217	91	295029	50.057	ug/l	99
83) tert-Butylbenzene	13.435	119	265994	48.341	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	347644	53.574	ug/l	100
85) sec-Butylbenzene	13.617	105	372684	51.919	ug/l	99
86) p-Isopropyltoluene	13.729	119	322242	53.032	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	188702	49.269	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	184771	45.917	ug/l	99
89) n-Butylbenzene	14.053	91	244714	49.455	ug/l	99
90) Hexachloroethane	14.335	117	64371	45.911	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	181781	47.744	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	24548	47.702	ug/l	96
93) 1,2,4-Trichlorobenzene	15.388	180	83568	43.652	ug/l	97
94) Hexachlorobutadiene	15.500	225	46500	41.866	ug/l	98
95) Naphthalene	15.641	128	249688	45.041	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	84562	46.550	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040125\
Data File : VN086191.D
Acq On : 01 Apr 2025 12:30
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 01:51:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040125\
Data File : VN086191.D
Acq On : 01 Apr 2025 12:30
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Quant Time: Apr 02 01:51:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 131862
Ion Ratio Lower Upper
168 100
99 61.7 49.4 74.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#2
Dichlorodifluoromethane
Concen: 60.662 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Tgt Ion: 85 Resp: 105974
Ion Ratio Lower Upper
85 100
87 34.0 16.5 49.5









#3
Chloromethane
Concen: 56.350 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 8630
Ion Ratio Lower Upper
50 100
52 31.5 26.0 39.0




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#4
Vinyl Chloride
Concen: 58.261 ug/l
RT: 2.506 min Scan# 94
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion: 62 Resp: 91233
Ion Ratio Lower Upper
62 100
64 32.3 25.2 37.8





#5
Bromomethane
Concen: 47.678 ug/l
RT: 2.936 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 50856
Ion Ratio Lower Upper
94 100
96 89.1 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#6
Chloroethane
Concen: 49.119 ug/l
RT: 3.106 min Scan# 196
Delta R.T. -0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion: 64 Resp: 48523
Ion Ratio Lower Upper
64 100
66 32.5 28.3 42.5





#7
 Trichlorofluoromethane
 Concen: 49.782 ug/l
 RT: 3.489 min Scan# 261
 Delta R.T. -0.000 min
 Lab File: VN086191.D
 Acq: 01 Apr 2025 12:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050



Tgt Ion:101 Resp: 140135
 Ion Ratio Lower Upper
 101 100
 103 65.7 51.7 77.5




Manual Integrations
 APPROVED


Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025



#8
 Diethyl Ether
 Concen: 42.596 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. -0.000 min
 Lab File: VN086191.D
 Acq: 01 Apr 2025 12:30

Tgt Ion: 74 Resp: 36707
 Ion Ratio Lower Upper
 74 100
 45 76.7 38.0 113.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.277 ug/l
RT: 4.365 min Scan# 410
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 70730
Ion Ratio Lower Upper
101 100
85 45.8 36.5 54.7
151 80.4 63.4 95.0




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#10
Methyl Iodide
Concen: 48.417 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion:142 Resp: 94384
Ion Ratio Lower Upper
142 100
127 52.1 37.5 56.3
141 14.9 11.0 16.6





#11
Tert butyl alcohol
Concen: 219.694 ug/l
RT: 5.524 min Scan# 607
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 59 Resp: 56164
Ion Ratio Lower Upper
59 100
57 10.6 7.9 11.9

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#12
1,1-Dichloroethene
Concen: 46.789 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion: 96 Resp: 63208
Ion Ratio Lower Upper
96 100
61 152.4 120.5 180.7
98 66.5 47.4 71.0





#13
Acrolein
Concen: 182.699 ug/l
RT: 4.177 min Scan# 378
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 49810
Ion Ratio Lower Upper
56 100
55 74.0 55.1 82.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#14
Allyl chloride
Concen: 40.294 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion: 41 Resp: 67909
Ion Ratio Lower Upper
41 100
39 76.9 55.5 83.3
76 52.2 39.1 58.7





#15
Acrylonitrile
Concen: 235.861 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 53 Resp: 159244
Ion Ratio Lower Upper
53 100
52 81.0 63.9 95.9
51 35.2 27.1 40.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#16
Acetone
Concen: 210.819 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30



Tgt Ion: 43 Resp: 116296
Ion Ratio Lower Upper
43 100
58 38.3 29.4 44.2









#17
Carbon Disulfide
Concen: 43.857 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 76 Resp: 194840
Ion Ratio Lower Upper
76 100
78 9.1 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#18
Methyl Acetate
Concen: 46.530 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion: 43 Resp: 63507
Ion Ratio Lower Upper
43 100
74 31.8 23.9 35.9





#19
Methyl tert-butyl Ether
Concen: 49.233 ug/l
RT: 5.794 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 73 Resp: 21794
Ion Ratio Lower Upper
73 100
57 20.2 16.1 24.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#20
Methylene Chloride
Concen: 48.030 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion: 84 Resp: 77510
Ion Ratio Lower Upper
84 100
49 100.1 83.3 124.9
51 29.6 24.8 37.2
86 63.5 48.7 73.1





#21
trans-1,2-Dichloroethene
Concen: 49.791 ug/l
RT: 5.788 min Scan# 651
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 73509
Ion Ratio Lower Upper
96 100
61 126.5 106.6 160.0
98 63.0 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#22
Diisopropyl ether
Concen: 47.921 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Tgt Ion: 45 Resp: 172108
Ion Ratio Lower Upper
45 100
43 52.4 41.2 61.8
87 36.2 28.5 42.7
59 12.9 11.2 16.8









#23
 Vinyl Acetate
 Concen: 241.172 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. -0.000 min
 Lab File: VN086191.D
 Acq: 01 Apr 2025 12:30

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tgt Ion: 43 Resp: 65660
 Ion Ratio Lower Upper
 43 100
 86 15.0 11.2 16.8




Manual Integrations
 APPROVED


Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025



#24
 1,1-Dichloroethane
 Concen: 48.404 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.000 min
 Lab File: VN086191.D
 Acq: 01 Apr 2025 12:30


Tgt Ion: 63 Resp: 127834
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.6 10.9
 100 4.8 2.5 7.5





#25
2-Butanone
Concen: 233.461 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 18583
Ion Ratio Lower Upper
43 100
72 34.5 27.4 41.0

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#26
2,2-Dichloropropane
Concen: 47.344 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion: 77 Resp: 120377
Ion Ratio Lower Upper
77 100
97 22.1 11.2 33.5





#27
 cis-1,2-Dichloroethene
 Concen: 51.224 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.006 min
 Lab File: VN086191.D
 Acq: 01 Apr 2025 12:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 96 Resp: 85938
 Ion Ratio Lower Upper
 96 100
 61 133.2 0.0 272.8
 98 63.2 0.0 124.8




Manual Integrations
 APPROVED


Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025



#28
 Bromochloromethane
 Concen: 48.515 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN086191.D
 Acq: 01 Apr 2025 12:30

Tgt Ion: 49 Resp: 54285
 Ion Ratio Lower Upper
 49 100
 129 2.8 0.0 4.2
 130 93.9 74.2 111.4





#29

Tetrahydrofuran

Concen: 253.767 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 124704





Ion	Ratio	Lower	Upper
42	100		
72	58.6	45.0	67.4
71	53.4	41.9	62.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 04/02/2025

Supervised By :Mahesh Dadoda 04/02/2025



#30

Chloroform

Concen: 51.821 ug/l

RT: 7.965 min Scan# 1022




Delta R.T. -0.000 min


Lab File: VN086191.D

Acq: 01 Apr 2025 12:30

Tgt Ion: 83 Resp: 151272

Ion	Ratio	Lower	Upper
83	100		
85	61.9	51.5	77.3





#31

Cyclohexane

Concen: 46.757 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. -0.000 min

Lab File: VN086191.D


Acq: 01 Apr 2025 12:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050



Tgt Ion: 56 Resp: 10519

Ion Ratio Lower Upper

56 100

69 30.9 28.0 42.0




84 90.8 75.0 112.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 04/02/2025

Supervised By :Mahesh Dadoda 04/02/2025



#32

1,1,1-Trichloroethane

Concen: 52.943 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30




Tgt Ion: 97 Resp: 146295


Ion Ratio Lower Upper

97 100

99 63.6 52.8 79.2

61 41.8 33.6 50.4





#33

1,2-Dichloroethane-d4

Concen: 54.185 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 97838

Ion Ratio Lower Upper

65 100





67 51.0 0.0 102.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 04/02/2025

Supervised By :Mahesh Dadoda 04/02/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30




Tgt Ion:114 Resp: 204993


Ion Ratio Lower Upper

114 100

63 19.1 0.0 39.6


88 16.7 0.0 32.6





#35
Dibromofluoromethane
Concen: 51.964 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 113 Resp: 7435
Ion Ratio Lower Upper
113 100
111 105.3 81.8 122.8
192 20.1 15.9 23.9

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#36
1,1-Dichloropropene
Concen: 53.617 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30



Tgt Ion: 75 Resp: 101016
Ion Ratio Lower Upper
75 100
110 36.2 17.8 53.3
77 30.8 25.0 37.4









#37
Ethyl Acetate
Concen: 48.680 ug/l
RT: 7.559 min Scan# 953
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 77375
Ion Ratio Lower Upper
43 100
61 16.2 13.0 19.4
70 14.2 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#38
Carbon Tetrachloride
Concen: 57.173 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Tgt Ion:117 Resp: 141628
Ion Ratio Lower Upper
117 100
119 95.3 77.0 115.4
121 29.8 25.6 38.4





#39

Methylcyclohexane

Concen: 51.425 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 96125
 Ion Ratio Lower Upper
 83 100
 55 75.2 54.2 81.2
 98 53.1 37.1 55.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 04/02/2025

Supervised By :Mahesh Dadoda 04/02/2025



#40

Benzene

Concen: 53.684 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30


Tgt Ion: 78 Resp: 317513
 Ion Ratio Lower Upper
 78 100
 77 21.5 18.9 28.3





#41
Methacrylonitrile
Concen: 54.310 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 41220
Ion Ratio Lower Upper
41 100
39 55.1 46.2 69.2
67 89.9 75.9 113.9
52 34.7 29.2 43.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#42
1,2-Dichloroethane
Concen: 55.093 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion: 62 Resp: 113538
Ion Ratio Lower Upper
62 100
98 8.9 0.0 18.6





#43
Isopropyl Acetate
Concen: 39.330 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 13546
Ion Ratio Lower Upper
43 100
61 30.2 25.0 37.6
87 16.2 13.0 19.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#44
Trichloroethene
Concen: 49.095 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Tgt Ion:130 Resp: 75257
Ion Ratio Lower Upper
130 100
95 90.8 0.0 191.6





#45

1,2-Dichloropropane

Concen: 52.691 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 7088

Ion Ratio Lower Upper

63 100





65 35.0 24.7 37.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 04/02/2025

Supervised By :Mahesh Dadoda 04/02/2025



#46

Dibromomethane

Concen: 55.405 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30




Tgt Ion: 93 Resp: 58614


Ion Ratio Lower Upper

93 100

95 81.0 66.2 99.2


174 99.0 80.1 120.1





#47
Bromodichloromethane
Concen: 55.345 ug/l
RT: 9.888 min Scan# 1348
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 122368
Ion Ratio Lower Upper
83 100
85 63.2 50.2 75.2
127 10.0 7.1 10.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#48
Methyl methacrylate
Concen: 54.205 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion: 41 Resp: 61608
Ion Ratio Lower Upper
41 100
69 109.4 83.1 124.7
39 57.1 45.6 68.4





#49
1,4-Dioxane
Concen: 1134.096 ug/l
RT: 9.694 min Scan# 1316
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 88 Resp: 3047
Ion Ratio Lower Upper
88 100
43 28.1 21.7 32.5
58 66.2 50.4 75.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#50
Toluene-d8
Concen: 54.542 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion: 98 Resp: 283232
Ion Ratio Lower Upper
98 100
100 64.7 53.1 79.7





#51
4-Methyl-2-Pentanone
Concen: 285.346 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 43344
Ion Ratio Lower Upper
43 100
58 45.9 37.4 56.0

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#52
Toluene
Concen: 58.421 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion: 92 Resp: 212069
Ion Ratio Lower Upper
92 100
91 169.6 136.2 204.4





#53
t-1,3-Dichloropropene
Concen: 53.936 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 75 Resp: 115814
Ion Ratio Lower Upper
75 100
77 34.4 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#54
cis-1,3-Dichloropropene
Concen: 53.797 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30



Tgt Ion: 75 Resp: 121877
Ion Ratio Lower Upper
75 100
77 31.2 25.3 37.9
39 39.2 29.6 44.4





#55

1,1,2-Trichloroethane

Concen: 56.188 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 97 Resp: 7784

Ion Ratio Lower Upper

97 100

83 84.8 71.2 106.8

85 52.5 45.4 68.0





99 64.0 49.5 74.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 04/02/2025

Supervised By :Mahesh Dadoda 04/02/2025



#56

Ethyl methacrylate

Concen: 51.295 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.001 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30




Tgt Ion: 69 Resp: 111441


Ion Ratio Lower Upper

69 100

41 52.5 41.6 62.4


39 29.6 21.8 32.8





#57
1,3-Dichloropropane
Concen: 56.160 ug/l
RT: 11.159 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 13215
Ion Ratio Lower Upper
76 100
78 32.3 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#58
2-Chloroethyl Vinyl ether
Concen: 304.296 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion: 63 Resp: 263133
Ion Ratio Lower Upper
63 100
106 31.7 25.0 37.6





#59
2-Hexanone
Concen: 290.158 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 321084
Ion Ratio Lower Upper
43 100
58 64.0 32.4 97.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#60
Dibromochloromethane
Concen: 57.261 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion:129 Resp: 102372
Ion Ratio Lower Upper
129 100
127 76.7 37.8 113.4





#61
1,2-Dibromoethane
Concen: 55.473 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:107 Resp: 7890
Ion Ratio Lower Upper
107 100
109 93.9 73.3 109.9

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#62
4-Bromofluorobenzene
Concen: 57.194 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion: 95 Resp: 105921
Ion Ratio Lower Upper
95 100
174 79.5 0.0 156.8
176 76.4 0.0 152.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 19589
Ion Ratio Lower Upper
117 100
82 56.8 46.7 70.1
119 33.4 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#64
Tetrachloroethene
Concen: 51.506 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion:164 Resp: 80488
Ion Ratio Lower Upper
164 100
166 126.4 103.7 155.5
129 100.3 77.1 115.7
131 92.4 73.3 109.9





#65
Chlorobenzene
Concen: 50.343 ug/l
RT: 11.888 min Scan# 112
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:112 Resp: 225670
Ion Ratio Lower Upper
112 100
114 31.0 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 52.796 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion:131 Resp: 87978
Ion Ratio Lower Upper
131 100
133 96.0 47.0 141.0
119 61.6 31.4 94.0





#67
Ethyl Benzene
Concen: 53.956 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 386894
Ion Ratio Lower Upper
91 100
106 31.2 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#68
m/p-Xylenes
Concen: 112.511 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Tgt Ion:106 Resp: 317096
Ion Ratio Lower Upper
106 100
91 207.5 163.9 245.9









#69
o-Xylene
Concen: 56.358 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 145782
Ion Ratio Lower Upper
106 100
91 218.0 109.1 327.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#70
Styrene
Concen: 59.027 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Tgt Ion:104 Resp: 260792
Ion Ratio Lower Upper
104 100
78 51.1 41.4 62.2
103 54.1 42.9 64.3









#71
Bromoform
Concen: 53.039 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Instrument : MSVOA_N
ClientSampleId : VSTDCCC050

Tgt Ion:173 Resp: 73884
Ion Ratio Lower Upper
173 100
175 48.3 24.3 73.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion:152 Resp: 111589
Ion Ratio Lower Upper
152 100
115 63.4 31.1 93.2
150 174.9 0.0 359.6





#73
Isopropylbenzene
Concen: 48.987 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:105 Resp: 37329
Ion Ratio Lower Upper
105 100
120 26.1 13.0 39.0

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#74
N-ethyl acetate
Concen: 46.828 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion: 43 Resp: 119388
Ion Ratio Lower Upper
43 100
70 54.9 43.0 64.4
55 32.4 23.8 35.8
61 28.6 23.4 35.0





#75
1,1,2,2-Tetrachloroethane
Concen: 44.496 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 11471
Ion Ratio Lower Upper
83 100
131 11.1 5.5 16.5
85 64.2 32.6 97.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#76
1,2,3-Trichloropropane
Concen: 45.386 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion: 75 Resp: 113386
Ion Ratio Lower Upper
75 100
77 179.3 83.7 251.0





#77
Bromobenzene
Concen: 47.181 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:156 Resp: 100250
Ion Ratio Lower Upper
156 100
77 202.8 104.1 312.3
158 98.5 49.0 147.0

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#78
n-propylbenzene
Concen: 51.446 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30



Tgt Ion: 91 Resp: 443917
Ion Ratio Lower Upper
91 100
120 23.3 11.7 35.0









#79
2-Chlorotoluene
Concen: 49.095 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 28138
Ion Ratio Lower Upper
91 100
126 32.8 16.4 49.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#80
1,3,5-Trimethylbenzene
Concen: 52.995 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion:105 Resp: 340956
Ion Ratio Lower Upper
105 100
120 47.2 23.7 71.1





#81
trans-1,4-Dichloro-2-butene
Concen: 42.843 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 40484
Ion Ratio Lower Upper
75 100
53 92.1 66.0 99.0
89 48.5 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#82
4-Chlorotoluene
Concen: 50.057 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30


Tgt Ion: 91 Resp: 295029
Ion Ratio Lower Upper
91 100
126 32.8 16.1 48.2





#83
tert-Butylbenzene
Concen: 48.341 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument : MSVOA_N
ClientSampleId : VSTDCCC050




Tgt Ion:119 Resp: 265994
Ion Ratio Lower Upper
119 100
91 69.4 34.3 102.9
134 27.0 13.5 40.5

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#84
1,2,4-Trimethylbenzene
Concen: 53.574 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30



Tgt Ion:105 Resp: 347644
Ion Ratio Lower Upper
105 100
120 44.1 21.9 65.8





#85

sec-Butylbenzene

Concen: 51.919 ug/l

RT: 13.617 min Scan# 1982

Delta R.T. 0.006 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 372684

Ion Ratio Lower Upper

105 100





134 20.4 10.1 30.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 04/02/2025

Supervised By :Mahesh Dadoda 04/02/2025



#86

p-Isopropyltoluene

Concen: 53.032 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30




Tgt Ion:119 Resp: 322242


Ion Ratio Lower Upper

119 100

134 25.5 12.6 37.8


91 25.6 12.3 36.8





#87
1,3-Dichlorobenzene
Concen: 49.269 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 188702
Ion Ratio Lower Upper
146 100
111 43.0 21.2 63.6
148 63.9 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#88
1,4-Dichlorobenzene
Concen: 45.917 ug/l
RT: 13.811 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Tgt Ion:146 Resp: 184771
Ion Ratio Lower Upper
146 100
111 41.8 20.4 61.3
148 64.8 32.1 96.5





#89

n-Butylbenzene

Concen: 49.455 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:	91	Resp:	244714
Ion	Ratio	Lower	Upper
91	100		
92	52.5	26.9	80.6
134	25.6	13.0	39.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 04/02/2025

Supervised By :Mahesh Dadoda 04/02/2025



#90

Hexachloroethane

Concen: 45.911 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. 0.006 min

Lab File: VN086191.D

Acq: 01 Apr 2025 12:30


Tgt Ion:	117	Resp:	64371
Ion	Ratio	Lower	Upper
117	100		
201	77.6	39.1	117.3





#91
1,2-Dichlorobenzene
Concen: 47.744 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 18178
Ion Ratio Lower Upper
146 100
111 44.4 22.3 66.8
148 63.6 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 47.702 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion: 75 Resp: 24548
Ion Ratio Lower Upper
75 100
155 79.9 41.0 123.0
157 101.6 53.7 161.1





#93
1,2,4-Trichlorobenzene
Concen: 43.652 ug/l
RT: 15.388 min Scan# 2284
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:180 Resp: 8356
Ion Ratio Lower Upper
180 100
182 96.1 47.0 141.0
145 33.2 15.5 46.5

Manual Integrations
APPROVED



Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025




#94
Hexachlorobutadiene
Concen: 41.866 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Tgt Ion:225 Resp: 46500
Ion Ratio Lower Upper
225 100
223 60.4 31.3 93.9
227 65.1 32.1 96.3





#95
Naphthalene
Concen: 45.041 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:128 Resp: 249688
Ion Ratio Lower Upper
128 100
127 13.2 10.2 15.4
129 10.8 9.0 13.6

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 46.550 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN086191.D
Acq: 01 Apr 2025 12:30

Tgt Ion:180 Resp: 84562
Ion Ratio Lower Upper
180 100
182 95.3 47.5 142.5
145 35.0 17.7 53.1

