

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040219\
 Data File : VN054892.D
 Acq On : 2 Apr 2019 11:00
 Operator : JC/SP
 Sample : VN0402WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0402WBL01

Quant Time: Apr 03 05:32:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	520942	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	847901	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	729388	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	248017	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	337038	47.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.98%	
35) Dibromofluoromethane	7.59	113	267457	47.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.62%	
50) Toluene-d8	10.09	98	1023021	48.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.36%	
62) 4-Bromofluorobenzene	12.40	95	307524	44.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040219\
Data File : VN054892.D
Acq On : 2 Apr 2019 11:00
Operator : JC/SP
Sample : VN0402WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0402WBL01

Quant Time: Apr 03 05:32:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

