

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040219\
 Data File : VN054894.D
 Acq On : 2 Apr 2019 11:53
 Operator : JC/SP
 Sample : K2186-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BELT-PARKWAY-STAGING-YARD-2

Quant Time: Apr 03 05:37:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	545493	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	885614	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	756488	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	255545	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	356050	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
35) Dibromofluoromethane	7.59	113	283356	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	10.09	98	1065190	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
62) 4-Bromofluorobenzene	12.40	95	323636	44.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.66%	
Target Compounds						
20) Methylene Chloride	4.55	84	17955	2.903	ug/l #	79
43) Isopropyl Acetate	8.18	43	14004	1.424	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040219\
Data File : VN054894.D
Acq On : 2 Apr 2019 11:53
Operator : JC/SP
Sample : K2186-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BELT-PARKWAY-STAGING-YARD-2

Quant Time: Apr 03 05:37:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

