

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040219\
 Data File : VN054896.D
 Acq On : 2 Apr 2019 12:46
 Operator : JC/SP
 Sample : K2186-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-TRACK-78-79

Quant Time: Apr 03 05:43:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	525142	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	871993	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	761747	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	260568	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	356354	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
35) Dibromofluoromethane	7.59	113	277585	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
50) Toluene-d8	10.09	98	1066973	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
62) 4-Bromofluorobenzene	12.40	95	321378	44.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.42%	
Target Compounds						
16) Acetone	3.83	43	17068	10.111	ug/l #	83
20) Methylene Chloride	4.55	84	17086	2.869	ug/l #	87
43) Isopropyl Acetate	8.17	43	14493	1.497	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040219\
Data File : VN054896.D
Acq On : 2 Apr 2019 12:46
Operator : JC/SP
Sample : K2186-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-TRACK-78-79

Quant Time: Apr 03 05:43:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

