

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040219\
 Data File : VN054900.D
 Acq On : 2 Apr 2019 14:31
 Operator : JC/SP
 Sample : PB118500TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB118500TB

Quant Time: Apr 03 05:55:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	538990	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	885223	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	776209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	277048	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	352821	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
35) Dibromofluoromethane	7.59	113	278372	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
50) Toluene-d8	10.09	98	1075789	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
62) 4-Bromofluorobenzene	12.40	95	336173	46.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.14%	
Target Compounds						
20) Methylene Chloride	4.55	84	9231	1.510	ug/l #	77
43) Isopropyl Acetate	8.17	43	15671	1.594	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040219\
Data File : VN054900.D
Acq On : 2 Apr 2019 14:31
Operator : JC/SP
Sample : PB118500TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB118500TB

Quant Time: Apr 03 05:55:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

