

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040219\
 Data File : VN054903.D
 Acq On : 2 Apr 2019 15:50
 Operator : JC/SP
 Sample : K2213-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S9B(8)

Quant Time: Apr 03 06:09:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	514312	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	856151	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	739007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	257757	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	340362	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
35) Dibromofluoromethane	7.59	113	271945	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
50) Toluene-d8	10.09	98	1034435	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
62) 4-Bromofluorobenzene	12.40	95	314762	44.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.20%	
Target Compounds						
16) Acetone	3.82	43	10777	6.519	ug/l	91
20) Methylene Chloride	4.54	84	12059	2.068	ug/l #	83
43) Isopropyl Acetate	8.17	43	13620	1.433	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040219\
Data File : VN054903.D
Acq On : 2 Apr 2019 15:50
Operator : JC/SP
Sample : K2213-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S9B(8)

Quant Time: Apr 03 06:09:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

