

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040219\
 Data File : VN054904.D
 Acq On : 2 Apr 2019 16:17
 Operator : JC/SP
 Sample : K2213-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S9A(6)

Quant Time: Apr 03 06:37:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	543230	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	904428	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	774031	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	266414	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	358769	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
35) Dibromofluoromethane	7.59	113	284593	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
50) Toluene-d8	10.09	98	1079044	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
62) 4-Bromofluorobenzene	12.40	95	327167	43.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.76%	
Target Compounds						
20) Methylene Chloride	4.55	84	11597	1.883	ug/l #	84
43) Isopropyl Acetate	8.17	43	14684	1.462	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040219\
Data File : VN054904.D
Acq On : 2 Apr 2019 16:17
Operator : JC/SP
Sample : K2213-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S9A(6)

Quant Time: Apr 03 06:37:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

