

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN040219\
 Data File : VN054891.D
 Acq On : 2 Apr 2019 10:34
 Operator : JC/SP
 Sample : VN0402MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0402MBL01

Quant Time: Apr 03 05:26:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	471473	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	753466	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	646128	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	198663	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	305110	48.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.02%	
35) Dibromofluoromethane	7.59	113	246619	49.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.20%	
50) Toluene-d8	10.09	98	925878	49.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.16%	
62) 4-Bromofluorobenzene	12.40	95	272088	43.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.62%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN040219\
Data File : VN054891.D
Acq On : 2 Apr 2019 10:34
Operator : JC/SP
Sample : VN0402MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0402MBL01

Quant Time: Apr 03 05:26:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.00 min

Lab File: VN054891.D

Acq: 2 Apr 2019 10:34

Instrument :

MSVOA_N

ClientSampled :


VN0402MBL01

Tot Ion:168 Resp: 471473

Ion Ratio Lower Upper

168 100

99 55.2 37.1 55.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.66

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 48.007 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN054891.D


Acq: 2 Apr 2019 10:34

Tgt Ion: 65 Resp: 305110

Ion Ratio Lower Upper

65 100

67 52.3 0.0 104.6



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0

8.03


150000


100000

50000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.58 min Scan# 2119

Delta R.T. -0.00 min

Lab File: VN054891.D

Acq: 2 Apr 2019 10:34

Instrument :

MSVOA_N

ClientSampleId :

VN0402MBL01


Tot Ion:114 Resp: 753466

Ion Ratio Lower Upper

114 100

63 22.8 0.0 35.2




88 15.7 0.0 30.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 49.095 ug/l

RT: 7.59 min Scan# 1809

Delta R.T. 0.00 min

Lab File: VN054891.D

Acq: 2 Apr 2019 10:34


Tgt Ion:113 Resp: 246619

Ion Ratio Lower Upper

113 100

111 102.4 82.2 123.4



192 19.9 19.2 28.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 49.577 ug/l
RT: 10.09 min Scan# 2586
Delta R.T. -0.00 min
Lab File: VN054891.D
Acq: 2 Apr 2019 10:34




Instrument :
MSVOA_N
ClientSampled :
VN0402MBL01


Tot Ion: 98 Resp: 925878
Ion Ratio Lower Upper
98 100
100 63.0 51.2 76.8



#62
4-Bromofluorobenzene
Concen: 43.807 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.00 min
Lab File: VN054891.D
Acq: 2 Apr 2019 10:34

Tgt Ion: 95 Resp: 272088
Ion Ratio Lower Upper
95 100
174 82.7 0.0 191.0
176 80.8 0.0 187.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN054891.D
Acq: 2 Apr 2019 10:34

Instrument :
MSVOA_N
ClientSampleId :
VN0402MBL01

Tot Ion:117 Resp: 646128
Ion Ratio Lower Upper
117 100
82 55.0 40.0 60.0
119 31.5 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN054891.D
Acq: 2 Apr 2019 10:34

Tgt Ion:152 Resp: 198663
Ion Ratio Lower Upper
152 100
115 57.3 27.2 81.6
150 154.8 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

