


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN040219\
Data File : VN054889.D
Acq On : 2 Apr 2019 8:34
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Title : SW846 8260
Last Update : Sat Mar 30 00:48:18 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	12022	PASS
75	95	30	60	51.2	29045	PASS
95	95	100	100	100.0	56674	PASS
96	95	5	9	6.5	3672	PASS
173	174	0.00	2	0.4	214	PASS
174	95	50	100	84.4	47826	PASS
175	174	5	9	7.1	3412	PASS
176	174	95	101	96.4	46112	PASS
177	176	5	9	6.4	2973	PASS