

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040220\
 Data File : VN060820.D
 Acq On : 2 Apr 2020 13:54
 Operator : JC/MD
 Sample : L2018-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Apr 03 08:41:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	202317	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	336901	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	312283	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	138180	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130746	47.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.16%	
35) Dibromofluoromethane	7.56	113	99569	48.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.76%	
50) Toluene-d8	10.08	98	413131	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
62) 4-Bromofluorobenzene	12.40	95	150784	49.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040220\
Data File : VN060820.D
Acq On : 2 Apr 2020 13:54
Operator : JC/MD
Sample : L2018-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Apr 03 08:41:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

