

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040220\
 Data File : VN060829.D
 Acq On : 2 Apr 2020 18:12
 Operator : JC/MD
 Sample : L2101-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 239-WC-S-SB-01

Quant Time: Apr 03 09:16:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	206315	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	343219	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	323916	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	143867	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	133863	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
35) Dibromofluoromethane	7.56	113	102431	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	10.08	98	426219	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
62) 4-Bromofluorobenzene	12.40	95	157284	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
Target Compounds						
16) Acetone	3.78	43	479959	478.411	ug/l	100
20) Methylene Chloride	4.52	84	6900	2.796	ug/l	98
43) Isopropyl Acetate	8.14	43	52591	8.880	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040220\
Data File : VN060829.D
Acq On : 2 Apr 2020 18:12
Operator : JC/MD
Sample : L2101-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
239-WC-S-SB-01

Quant Time: Apr 03 09:16:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

