

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040220\
 Data File : VN060832.D
 Acq On : 2 Apr 2020 19:20
 Operator : JC/MD
 Sample : L2103-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-60-61-COMP

Quant Time: Apr 03 09:22:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	199194	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	329354	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	309689	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	141467	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	128379	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
35) Dibromofluoromethane	7.57	113	98769	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	10.08	98	408993	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
62) 4-Bromofluorobenzene	12.40	95	151722	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
Target Compounds						
16) Acetone	3.78	43	6746	6.965	ug/l	Qvalue 100
20) Methylene Chloride	4.52	84	4007	1.682	ug/l	95
43) Isopropyl Acetate	8.14	43	51118	8.995	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040220\
Data File : VN060832.D
Acq On : 2 Apr 2020 19:20
Operator : JC/MD
Sample : L2103-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-60-61-COMP

Quant Time: Apr 03 09:22:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

