

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040220\
 Data File : VN060834.D
 Acq On : 2 Apr 2020 20:05
 Operator : JC/MD
 Sample : L2104-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BUS-WASH

Quant Time: Apr 03 09:28:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	196483	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	326184	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	310928	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	149366	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	127176	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
35) Dibromofluoromethane	7.56	113	97380	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
50) Toluene-d8	10.08	98	408258	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
62) 4-Bromofluorobenzene	12.40	95	155111	52.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.08%	
Target Compounds						
6) Chloroethane	2.69	64	5326	3.329	ug/l	98
16) Acetone	3.78	43	43440	45.467	ug/l	97
20) Methylene Chloride	4.53	84	9697	4.126	ug/l	98
25) 2-Butanone	6.80	43	20990	13.577	ug/l	98
43) Isopropyl Acetate	8.14	43	49790	8.846	ug/l #	91
51) 4-Methyl-2-Pentanone	9.97	43	243621	76.065	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040220\
Data File : VN060834.D
Acq On : 2 Apr 2020 20:05
Operator : JC/MD
Sample : L2104-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BUS-WASH

Quant Time: Apr 03 09:28:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

