

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040220\
 Data File : VN060837.D
 Acq On : 2 Apr 2020 21:13
 Operator : JC/MD
 Sample : L2122-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-01-040120

Quant Time: Apr 03 09:35:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	190592	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	323069	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	303159	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	141351	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	125588	48.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.00%	
35) Dibromofluoromethane	7.57	113	96226	49.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.52%	
50) Toluene-d8	10.08	98	398426	49.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.86%	
62) 4-Bromofluorobenzene	12.40	95	149283	51.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.10%	
Target Compounds						
16) Acetone	3.79	43	8659	9.343	ug/l	93
20) Methylene Chloride	4.53	84	4110	1.803	ug/l	90
43) Isopropyl Acetate	8.14	43	59581	10.688	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040220\
Data File : VN060837.D
Acq On : 2 Apr 2020 21:13
Operator : JC/MD
Sample : L2122-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-01-040120

Quant Time: Apr 03 09:35:56 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

