

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040221\
 Data File : VN066475.D
 Acq On : 2 Apr 2021 18:15
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 03 03:25:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 01 14:08:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	368068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.510	114	547519	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	522806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	213301	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.944	65	226066	46.20	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.40%
35) Dibromofluoromethane	7.504	113	180568	49.75	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.50%
50) Toluene-d8	10.025	98	679694	49.37	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.74%
62) 4-Bromofluorobenzene	12.345	95	204553	42.78	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.56%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040221\
Data File : VN066475.D
Acq On : 2 Apr 2021 18:15
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 03 03:25:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 01 14:08:22 2021
Response via : Initial Calibration

