

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040221\
 Data File : VN066476.D
 Acq On : 2 Apr 2021 18:38
 Operator : JC/MD
 Sample : M1910-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-4121

Quant Time: Apr 03 03:25:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 01 14:08:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	338668	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	503983	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	475903	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	222290	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.944	65	212447	47.19	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.38%
35) Dibromofluoromethane	7.504	113	168125	50.32	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.64%
50) Toluene-d8	10.025	98	625186	49.33	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.66%
62) 4-Bromofluorobenzene	12.345	95	196071	44.54	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.08%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040221\
Data File : VN066476.D
Acq On : 2 Apr 2021 18:38
Operator : JC/MD
Sample : M1910-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-4121

Quant Time: Apr 03 03:25:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 01 14:08:22 2021
Response via : Initial Calibration

