

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040221\
 Data File : VN066479.D
 Acq On : 2 Apr 2021 19:49
 Operator : JC/MD
 Sample : PB135489TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135489TB

Quant Time: Apr 03 03:26:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 01 14:08:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	360183	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.510	114	535758	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	476385	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	209656	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.944	65	223689	46.72	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.44%
35) Dibromofluoromethane	7.504	113	181694	51.16	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.32%
50) Toluene-d8	10.025	98	647377	48.05	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.10%
62) 4-Bromofluorobenzene	12.345	95	186686	39.90	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.80%#
Target Compounds						
					Qvalue	
16) Acetone	3.736	43	99426	38.43	ug/l	99
20) Methylene Chloride	4.441	84	13514	1.66	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040221\
Data File : VN066479.D
Acq On : 2 Apr 2021 19:49
Operator : JC/MD
Sample : PB135489TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135489TB

Quant Time: Apr 03 03:26:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 01 14:08:22 2021
Response via : Initial Calibration

