

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040222\
 Data File : VN071809.D
 Acq On : 02 Apr 2022 19:46
 Operator : JC\MD
 Sample : N2200-21
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Apr 04 01:09:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	643768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1105015	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1109742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	406368	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	366167	49.823	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.640%
35) Dibromofluoromethane	8.016	113	316762	48.715	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.420%
50) Toluene-d8	10.439	98	1388424	52.197	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.400%
62) 4-Bromofluorobenzene	12.727	95	457056	53.129	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.260%
Target Compounds						
16) Acetone	4.287	43	6859	2.184	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040222\
Data File : VN071809.D
Acq On : 02 Apr 2022 19:46
Operator : JC\MD
Sample : N2200-21
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Apr 04 01:09:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

