

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040224\
 Data File : VN081650.D
 Acq On : 02 Apr 2024 12:26
 Operator : JC\MD
 Sample : VN0402WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0402WBL01

Quant Time: Apr 03 01:31:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	318617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	596075	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	541101	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	215284	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	228661	56.276	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.560%
35) Dibromofluoromethane	8.171	113	196446	52.221	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.440%
50) Toluene-d8	10.571	98	709907	52.204	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.400%
62) 4-Bromofluorobenzene	12.847	95	223278	46.598	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	93.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040224\
Data File : VN081650.D
Acq On : 02 Apr 2024 12:26
Operator : JC\MD
Sample : VN0402WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0402WBL01

Quant Time: Apr 03 01:31:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

