

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040224\
 Data File : VN081654.D
 Acq On : 02 Apr 2024 15:24
 Operator : JC\MD
 Sample : P1927-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11991

Quant Time: Apr 03 01:33:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	350180	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	677171	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	648559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	256673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	247030	55.317	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.640%
35) Dibromofluoromethane	8.165	113	216271	50.606	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.220%
50) Toluene-d8	10.565	98	840658	54.416	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.840%
62) 4-Bromofluorobenzene	12.847	95	264396	48.571	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	97.140%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	36920	25.582	ug/l	96
20) Methylene Chloride	5.277	84	16040	3.393	ug/l #	87
43) Isopropyl Acetate	8.688	43	239459	23.457	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040224\
Data File : VN081654.D
Acq On : 02 Apr 2024 15:24
Operator : JC\MD
Sample : P1927-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11991

Quant Time: Apr 03 01:33:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

