

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040225\
 Data File : VN086211.D
 Acq On : 02 Apr 2025 16:19
 Operator : JC\MD
 Sample : Q1693-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WCS-TP-2

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/03/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Apr 03 01:36:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	119883	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	201489	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	191161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	86961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	89273	54.382	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.760%			
35) Dibromofluoromethane	8.165	113	67430	47.946	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.900%			
50) Toluene-d8	10.565	98	251472	49.268	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.540%			
62) 4-Bromofluorobenzene	12.847	95	98084	53.884	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 107.760%			
Target Compounds					Qvalue	
16) Acetone	4.418	43	24143	48.139	ug/l	91
20) Methylene Chloride	5.271	84	8812	6.006	ug/l #	83
43) Isopropyl Acetate	8.676	43	10094m	2.982	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040225\
Data File : VN086211.D
Acq On : 02 Apr 2025 16:19
Operator : JC\MD
Sample : Q1693-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WCS-TP-2

Quant Time: Apr 03 01:36:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/03/2025
Supervised By :Mahesh Dadoda 04/03/2025

