

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040319\
 Data File : VN054925.D
 Acq On : 3 Apr 2019 14:41
 Operator : JC/SP
 Sample : K2203-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Apr 04 01:11:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	509909	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	871726	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	750455	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	275819	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	342525	49.83	ug/l	0.00
Spiked Amount			Recovery	=	99.66%	
35) Dibromofluoromethane	7.59	113	277245	47.70	ug/l	0.00
Spiked Amount			Recovery	=	95.40%	
50) Toluene-d8	10.09	98	1042648	48.26	ug/l	0.00
Spiked Amount			Recovery	=	96.52%	
62) 4-Bromofluorobenzene	12.40	95	321388	44.72	ug/l	0.00
Spiked Amount			Recovery	=	89.44%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.58	94	158	Below Cal	#	2
13) Acrolein	3.61	56	143	0.447	ug/l #	67
16) Acetone	3.82	43	8786	5.360	ug/l	89
18) Methyl Acetate	4.33	43	1731	Below Cal	#	89
20) Methylene Chloride	4.54	84	7256	1.255	ug/l	97
29) Tetrahydrofuran	7.22	42	683	0.406	ug/l #	30
30) Chloroform	7.38	83	4470	0.450	ug/l	95
31) Cyclohexane	7.66	56	11049	Below Cal	#	19
36) 1,1-Dichloropropene	7.67	75	38833	4.781	ug/l #	48
38) Carbon Tetrachloride	7.66	117	49150	6.230	ug/l #	16
43) Isopropyl Acetate	8.17	43	16074	1.661	ug/l #	94
48) Methyl methacrylate	9.18	41	7958	1.517	ug/l #	11
49) 1,4-Dioxane	9.27	88	112	2.151	ug/l #	1
51) 4-Methyl-2-Pentanone	10.03	43	10162	1.867	ug/l #	36
54) cis-1,3-Dichloropropene	9.90	75	59306	6.812	ug/l #	70
57) 1,3-Dichloropropane	10.76	76	10349	1.171	ug/l #	43
59) 2-Hexanone	10.76	43	172320	49.467	ug/l #	47
71) Bromoform	12.40	173	3192	0.991	ug/l #	1
74) N-amyl acetate	11.87	43	17365	2.472	ug/l #	51
75) 1,1,2,2-Tetrachloroethane	12.41	83	159	Below Cal	#	26
76) 1,2,3-Trichloropropane	12.40	75	165767	34.595	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	4973	0.269	ug/l	92
81) trans-1,4-Dichloro-2-buten	12.40	75	165767	134.567	ug/l #	11
83) tert-Butylbenzene	13.03	119	621	Below Cal	#	78
84) 1,2,4-Trimethylbenzene	13.04	105	4412	0.250	ug/l	95
95) Naphthalene	15.14	128	3130	0.315	ug/l #	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040319\
Data File : VN054925.D
Acq On : 3 Apr 2019 14:41
Operator : JC/SP
Sample : K2203-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Apr 04 01:11:56 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41


Instrument :
MSVOA_N
ClientSampled :

Tot Ion:168 Resp: 509909

Ion Ratio Lower Upper

168 100

99 54.5 37.1 55.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

200000

150000

100000



50000

0

7.60

7.70

Time-->



#5

Bromomethane

Concen: Below Cal

RT: 2.58 min Scan# 251

Delta R.T. 0.01 min

Lab File: VN054925.D


Acq: 3 Apr 2019 14:41

Tgt Ion: 94 Resp: 158

Ion Ratio Lower Upper

94 100

96 0.0 77.0 115.4#



Abundance Ion 93.90 (93.60 to 94.60): VNO

Ion 95.90 (95.60 to 96.60): VNO

2.58

200

150

100

50

0

2.56


2.57


2.58

2.59

2.60

Time-->





#13

Acrolein

Concen: 0.447 ug/l

RT: 3.61 min Scan# 570

Delta R.T. 0.00 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 56 Resp: 143

Ion Ratio Lower Upper

56 100

55 97.9 56.7 85.1#



#16

Acetone

Concen: 5.360 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.00 min

Lab File: VN054925.D




Acq: 3 Apr 2019 14:41


Tgt Ion: 43 Resp: 8786

Ion Ratio Lower Upper

43 100

58 30.7 29.9 44.9









#18
Methyl Acetate
Concen: Below Cal
RT: 4.33 min Scan# 795
Delta R.T. -0.00 min
Lab File: VN054925.D
Acq: 3 Apr 2019 14:41




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 1731
Ion Ratio Lower Upper
43 100
74 31.8 20.8 31.2#



#20
Methylene Chloride
Concen: 1.255 ug/l
RT: 4.54 min Scan# 860
Delta R.T. -0.01 min
Lab File: VN054925.D
Acq: 3 Apr 2019 14:41

Tgt Ion: 84 Resp: 7256
Ion Ratio Lower Upper
84 100
49 114.8 89.7 134.5
51 37.9 28.2 42.4
86 64.9 50.8 76.2





#29

Tetrahydrofuran

Concen: 0.406 ug/l

RT: 7.22 min Scan# 1693

Delta R.T. -0.00 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41

Instrument :

MSVOA_N

ClientSampled :





Tot Ion: 42 Resp: 683

Ion Ratio Lower Upper

42 100

72 0.0 37.8 56.6#

71 0.0 36.2 54.2#



#30

Chloroform

Concen: 0.450 ug/l

RT: 7.38 min Scan# 1743

Delta R.T. 0.00 min

Lab File: VN054925.D




Acq: 3 Apr 2019 14:41


Tgt Ion: 83 Resp: 4470

Ion Ratio Lower Upper

83 100

85 61.6 52.2 78.4





#31

Cyclohexane

Concen: Below Cal

RT: 7.66 min Scan# 1831

Delta R.T. 0.01 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 56 Resp: 11049

Ion Ratio Lower Upper

56 100

69 156.1 27.8 41.6#




84 124.6 74.6 111.8#



Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 84.00 (83.70 to 84.70): VN0



#33

1,2-Dichloroethane-d4

Concen: 49.831 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN054925.D


Acq: 3 Apr 2019 14:41

Tgt Ion: 65 Resp: 342525

Ion Ratio Lower Upper



65 100


67 53.2 0.0 104.6



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN054925.D
 Acq: 3 Apr 2019 14:41

Instrument :
 MSVOA_N
 ClientSampled :

Tot Ion:114 Resp: 871726
 Ion Ratio Lower Upper
 114 100
 63 21.9 0.0 35.2
 88 15.9 0.0 30.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 47.705 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN054925.D
 Acq: 3 Apr 2019 14:41

Tgt Ion:113 Resp: 277245
 Ion Ratio Lower Upper
 113 100
 111 101.2 82.2 123.4
 192 21.0 19.2 28.8



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.781 ug/l
 RT: 7.67 min Scan# 1833
 Delta R.T. -0.12 min
 Lab File: VN054925.D
 Acq: 3 Apr 2019 14:41




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 75 Resp: 38833
 Ion Ratio Lower Upper
 75 100
 110 8.9 19.1 57.5#
 77 0.0 25.0 37.4#



#38
 Carbon Tetrachloride
 Concen: 6.230 ug/l
 RT: 7.66 min Scan# 1832
 Delta R.T. -0.11 min
 Lab File: VN054925.D
 Acq: 3 Apr 2019 14:41

Tgt Ion: 117 Resp: 49150
 Ion Ratio Lower Upper
 117 100
 119 5.2 76.3 114.5#
 121 0.0 24.6 36.8#





#43

Isopropyl Acetate

Concen: 1.661 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.00 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 43 Resp: 16074

Ion Ratio Lower Upper




43 100

61 18.6 15.8 23.8

87 10.3 12.0 18.0#



Abundance Ion 43.00 (42.70 to 43.70): VNC
Ion 61.00 (60.70 to 61.70): VNC
Ion 87.00 (86.70 to 87.70): VNC



#48

Methyl methacrylate

Concen: 1.517 ug/l

RT: 9.18 min Scan# 2304

Delta R.T. -0.02 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41


Tgt Ion: 41 Resp: 7958

Ion Ratio Lower Upper



41 100


69 0.0 72.6 108.8#

39 0.0 47.5 71.3#



Abundance Ion 41.00 (40.70 to 41.70): VNC
Ion 69.00 (68.70 to 69.70): VNC
Ion 39.00 (38.70 to 39.70): VNC





#49

1,4-Dioxane

Concen: 2.151 ug/l

RT: 9.27 min Scan# 2331

Delta R.T. 0.06 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41

Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 88 Resp: 112

Ion Ratio Lower Upper




88 100

43 1206657.1 24.1 36.1#

58 5545.5 50.8 76.2#



Abundance Ion 88.00 (87.70 to 88.70): VN0
Ion 43.00 (42.70 to 43.70): VN0
Ion 58.00 (57.70 to 58.70): VN0



#50

Toluene-d8

Concen: 48.256 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN054925.D


Acq: 3 Apr 2019 14:41

Tgt Ion: 98 Resp: 1042648



Ion Ratio Lower Upper


98 100

100 63.5 51.2 76.8



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0









#51
 4-Methyl-2-Pentanone
 Concen: 1.867 ug/l
 RT: 10.03 min Scan# 2568
 Delta R.T. 0.04 min
 Lab File: VN054925.D
 Acq: 3 Apr 2019 14:41




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 43 Resp: 10162
 Ion Ratio Lower Upper
 43 100
 58 0.0 31.8 47.6#



#54
 cis-1,3-Dichloropropene
 Concen: 6.812 ug/l
 RT: 9.90 min Scan# 2529
 Delta R.T. 0.06 min
 Lab File: VN054925.D
 Acq: 3 Apr 2019 14:41

Tgt Ion: 75 Resp: 59306
 Ion Ratio Lower Upper
 75 100
 77 0.8 25.4 38.0#
 39 37.8 37.2 55.8









#57
 1,3-Dichloropropane
 Concen: 1.171 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.05 min
 Lab File: VN054925.D
 Acq: 3 Apr 2019 14:41




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 76 Resp: 10349
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.7 38.5#



#59
 2-Hexanone
 Concen: 49.467 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.01 min
 Lab File: VN054925.D
 Acq: 3 Apr 2019 14:41

Tgt Ion: 43 Resp: 172320
 Ion Ratio Lower Upper
 43 100
 58 17.4 28.0 83.9#





#62

4-Bromofluorobenzene

Concen: 44.725 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 95 Resp: 321388

Ion Ratio Lower Upper

95 100

174 85.4 0.0 191.0




176 82.4 0.0 187.4



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): VN0

Ion 175.80 (175.50 to 176.50): VN0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. -0.00 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41


Tgt Ion: 117 Resp: 750455

Ion Ratio Lower Upper

117 100

82 55.0 40.0 60.0



119 31.7 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): VN0

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): VN0









#71
Bromoform
Concen: 0.991 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.27 min
Lab File: VN054925.D
Acq: 3 Apr 2019 14:41




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion:173 Resp: 3192
Ion Ratio Lower Upper
173 100
175 587.6 24.4 73.2#
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN054925.D
Acq: 3 Apr 2019 14:41

Tgt Ion:152 Resp: 275819
Ion Ratio Lower Upper
152 100
115 57.4 27.2 81.6
150 154.3 0.0 349.2





#74

N-amvl acetate

Concen: 2.472 ug/l

RT: 11.87 min Scan# 3141

Delta R.T. -0.20 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41

Instrument :

MSVOA_N

ClientSampleId :

Tot Ion: 43 Resp: 17365


Ion Ratio Lower Upper

43 100

70 2.4 35.4 53.0#

55 5.5 19.8 29.6#

61 8.9 19.8 29.6#






Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0



#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.41 min Scan# 3307

Delta R.T. -0.10 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41


Tgt Ion: 83 Resp: 159

Ion Ratio Lower Upper

83 100

131 0.0 5.8 17.4#



85 0.0 32.2 96.6#




Abundance Ion 82.90 (82.60 to 83.60): VN0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VN0





#76

1,2,3-Trichloropropane

Concen: 34.595 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 75 Resp: 165767

Ion Ratio Lower Upper

75 100

77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0

12.40

100000

80000



60000

40000

20000

0

Time--> 12.30 12.40 12.50



#80

1,3,5-Trimethylbenzene

Concen: 0.269 ug/l

RT: 12.73 min Scan# 3408

Delta R.T. 0.00 min

Lab File: VN054925.D


Acq: 3 Apr 2019 14:41

Tgt Ion:105 Resp: 4973

Ion Ratio Lower Upper

105 100

120 44.3 25.0 75.0



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

12.73

4000


3000


2000

1000

0

Time--> 12.70 12.75









#81
trans-1,4-Dichloro-2-butene
Concen: 134.567 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.10 min
Lab File: VN054925.D
Acq: 3 Apr 2019 14:41




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 75 Resp: 165767
Ion Ratio Lower Upper
75 100
53 0.0 76.2 114.2#
89 0.0 39.2 58.8#



#83
tert-Butylbenzene
Concen: Below Cal
RT: 13.03 min Scan# 3502
Delta R.T. 0.04 min
Lab File: VN054925.D
Acq: 3 Apr 2019 14:41

Tgt Ion: 119 Resp: 621
Ion Ratio Lower Upper
119 100
91 69.4 30.9 92.8
134 0.0 13.2 39.6#





#84

1,2,4-Trimethylbenzene

Concen: 0.250 ug/l

RT: 13.04 min Scan# 3503

Delta R.T. -0.00 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:105 Resp: 4412

Ion Ratio Lower Upper




105 100

120 43.4 23.4 70.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#95

Naphthalene

Concen: 0.315 ug/l

RT: 15.14 min Scan# 4156

Delta R.T. 0.00 min

Lab File: VN054925.D

Acq: 3 Apr 2019 14:41


Tgt Ion:128 Resp: 3130

Ion Ratio Lower Upper

128 100

127 10.8 10.1 15.1

129 4.5 8.7 13.1#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

