

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040319\
 Data File : VN054916.D
 Acq On : 3 Apr 2019 10:44
 Operator : JC/SP
 Sample : VN0403WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0403WBL01

Quant Time: Apr 04 01:47:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	510155	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	850987	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	720847	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	246139	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	338031	49.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.30%	
35) Dibromofluoromethane	7.59	113	268809	47.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.76%	
50) Toluene-d8	10.09	98	1021663	48.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.88%	
62) 4-Bromofluorobenzene	12.40	95	310974	44.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.66%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040319\
Data File : VN054916.D
Acq On : 3 Apr 2019 10:44
Operator : JC/SP
Sample : VN0403WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0403WBL01

Quant Time: Apr 04 01:47:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

