

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040319\
 Data File : VN054922.D
 Acq On : 3 Apr 2019 13:22
 Operator : JC/SP
 Sample : K2221-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S2B(10)

Quant Time: Apr 04 02:14:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	508309	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	856493	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	722714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	247925	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	336075	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
35) Dibromofluoromethane	7.59	113	274323	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
50) Toluene-d8	10.09	98	1030867	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
62) 4-Bromofluorobenzene	12.40	95	305049	43.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.42%	
Target Compounds						
16) Acetone	3.82	43	17710	10.838	ug/l #	78
20) Methylene Chloride	4.55	84	63422	11.004	ug/l #	79
43) Isopropyl Acetate	8.17	43	15893	1.671	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040319\
Data File : VN054922.D
Acq On : 3 Apr 2019 13:22
Operator : JC/SP
Sample : K2221-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S2B(10)

Quant Time: Apr 04 02:14:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

